FiO_2 CROSS REFERENCE CHART

	NO SENSOR REQUIRED	MAX-1 R113P90	MAX-2 R107P20	MAX-3 R109P11	MAX-7 R109P10	MAX-8 R107P22	MAX-9 R114P80	MAX-10 R112P10	MAX-11 R113P10	MAX-11i R113P06	MAX-12 R109P45	MAX-12A R109P48	MAX-13 R115P10	MAX-13-250 R125P07	MAX-14 R116P80	MAX-16 R114P70	MAX-17 R116P10	MAX-18 R116P40	MAX-19 R116P60	MAX-22cc R116P30	MAX-43 R109P08	MAX-250 R125P01-002	MAX-250E R125P03-002	MAX-250+ R125P02-011
	12 month warranty	*18 month warranty*	*12 month warranty*	*12 month warranty*	*18 month warranty*	*14 month warranty*	*12 month warranty*	*12 month warranty*	*12 month warranty*	*14 month warranty*	*14 month warranty*	*14 month warranty*	*24 month warranty*	*14 month warranty*	*24 month warranty*	*12 month warranty*	*12 month warranty*	*12 month warranty*	*12 month warranty*	*12 month warranty*	*24 month warranty*	*24 month warranty*	*24 month warranty*	
Airshields (DRÄGER)							6736140					8362030, C2000 Isolette Incubator	6735142	6735142											
Analytical Industries		PSR-11-915	PSR-11-915-1			PSR-11-915-3	PSR-11-55	PSR-11-915-4	PSR-11-915-2	PSR-11-915-2	PSR-11-917-M		PSR-11-917-J	PSR-11-917-J		PSR-11-917-M	PSR-11-77	PSR-11-33	PSR-11-33			PSR-11-75-KE1	PSR-11-917-J1	
Bio Med Devices (BMD) (BMD #4433 "Cross Vent" use MAX-13C)													10006, M-10/M-2 Monitors	10006, M-10/M-2 Monitors					65004					
City Technology				MOX-3					MOX 16		MOX-9		MOX 2				MOX-4							
Criticare (CSI)																								
Datascope											Multinex Analyzer, 0600-00-0002 (use MAX-12F)		Anestar S	Anestar S										
GE Medical (DATEX•OHMEDA)	5410, 5420, ADU Models	0237-2034-700, Excel Modulus, 4700 Oxicap, 5250 RGM, 5120/5125 Monitor, 7800				0237-2034-700, Excel Modulus, 5250 RGM (ext. only), 5120/5125 Monitor		6050-0004-110, Modulus SE, 7900 Series, Aestiva Series, SmartVent S-5, Aespire, Excel SE													Giraffe			
Devilbiss															02ANA-101									
Dräger (No sensor available for "Julian" models # 6850930)			Narkomed Series, Fabias Series, 6400, GS Series, Trio Series, 6850645, 6803290						Narkomed Series, Fabias Series, 6400, Savina, GS Series, Trio Series, 6803290	Evita 2, Evita 4, Baby Log, I8000, 6850645	Infinity Delta													
Hamilton Medical		HM-10							HM-14, HM-13		HM-12, 396008, 396009, Galileo, Raphael, Arabella, Aladdin, 51254		HM-11	HM-11			HM-03		396000, 396002,, 51250, HM-01, Veolar, Amadeus, Arabella					
HP									15201A															
Hudson RCI (TELEFLEX) (No sensor available for Hudson 5803 & 5804)		5556					5500, V-05, 5590, 5577, 5584, 5595, 6477		5557, V-14		5566, V-12		5567, V-11	5567			5558, V-03	5569, V-02, 5525, 5528, 5529, 5550, 5552, 5568	5569, 5568, C1, V-01					
IMT Medical											PF300													
Invivo			Magnitude						Magnitude				9445	9445										
Maxtec																						OM-25A	OM-25ME, OM-25AE, MAXO2+ AE, MaxBlend	MAXO2+A
Mercury Medical		10-103-07	10-103-12			10-103-13	10-103-03	10-103-10	10-103-07		10-103-06		10-103-05	10-103-05	10-103-13		10-103-02	10-103-01	10-103-00			10-103-11	10-103-08	
MSA													406931 MiniOx I, II, III, MiniOx 3000	406931 MiniOx I, II, III, MiniOx 3000	806572, MiniOx 1A									
Newport Medical																	0M105 (OM100) Wave Vent, E200						SEN-1800P, E-500	
Pacifitech		PT-14A	PT-24A		PT-03	PT-28A	PT-05	PT-20A	PT-14A		PT-12A		PT-11A	PT-11A	PT-26		PT-03	PT-02	PT-01			PT-250	PT-25A	
Penlon																	MTS							
Purtian-Bennett (MALLINKRODT)													4-020933-00 (7820 Monitor) 7200 Ventilator	4-020933-00 (7820 Monitor) 7200 Ventilator		4-072214-00 840, 740, 760 & PB2000 Tester Vents	(Note: early-model PB 840 vents use MAX-250 sensor) (Note: early-model PB 740 vents use MAX-25 sensor)					
Respironics													Espirit Ventilator, Lifecare, (#PLV-102) Vision, 37-079, 1001454	Espirit Ventilator, Lifecare, (#PLV-102) Vision, 37-079, 1001454										
Sechrist Industries																	43170, 650 Vent							
Siemens (No sensor available for #64 86 679 E4070E)											SV300, 710, SV900, C/D, 9000, 9004979-E347E, 6419332-E380E													
Spacelabs											90513, Capnograph											Sirius		
Teledyne		R15, #C-41131	R31MED		AX300, MX300	R32MED	R13, #B-39636		R23, #C-61339		R24, #B-62321		R17MED, #C-43690	R17MED, #C-43690			T7, A-51327, 60T, 191T, 200T	C1R, C2R, #B-39022	T1, #A-37016	R22MED, #C-44611		R291MED	R29MED	
Versamed																Versamed Vent								
Viasys (EME & BIRD)	Cardiopulminary Pulmonetic										Avea Ventilator		Bird 6400/8400, VIP Vent, 66030	Bird 6400/8400, VIP Vent, 66030								Vela, 15974 *(use Max-250K)*	Sentry Blender, 66045	
VTI		101272, 100352	101372			101272			101372, 100352		103100		100580	100580										

Blue Text = Part Number
Black Text = Model Number

Airshields, Analytical Industries, Bio Med Devices, City Technology, Criticare, CSI, Datascope, GE Medical, Datex•Ohmeda, Devilbiss, Dräger, Hamilton Medical, HP, Hudson RCI, Teleflex, IMT Medical, Invivo, Maxtec, Mercury Medical,MSA, Newport Medical, Pacifitech, Penlon, Puritan-Bennett, Respironics, Sechrist Industries, Siemens, Spacelabs, Teledyne, Versamed, Viasys, EME, Bird, and VTI are trademarks of their respective corporations. Maxtec® is not affiliated with nor endorsed by these companies.

ML165 Rev. E (#1650)

Flow Sensors

SpiroQuant A
[R122P01]

SpiroQuant H
[R122P04]

Serial Number Definition Chart

Maxtec® Oxygen Sensor serial numbers are located on the label. We include the lot number, job number, line item number and sequence number all to identify when the sensor was manufactured. The model number is listed separately on the label.



Maxtec® Frequent Buyer Program

Ask a Maxtec® representative today about how to qualify for a free sensor.